


Model	Z 040 K3N GSG 125
Serial Number	B2353

Calibration Coefficient Using Calibration Substrate CSR-1		
OPEN	C = 12.2 fF	Offset length = 186 µm
SHORT	L = 19.3 pH	Offset length = 115 µm
LOAD (for Anritsu)	L = -9.9 pH	
LOAD (for Agilent)	Delay = 0.0198 ps	Line: 5 Ohms

Uncalibrated Performance @ 40 GHz :	
Return Loss Into A 50 Ω LOAD :	-20,0 dB
Maximum Insertion Loss:	-0,67 dB




Return Loss

Insertion Loss

The IZI PROBE is developed in conjunction with
ROSENBERGER HOCHFREQUENZTECHNIK
GmbH & Co., Germany

Rosenberger®



www.suss.com

SDSS MicroTec